

Pathology Lab Kits & Blood Bank kit
2016-17

s.#.	FIRMS NAME	REMARKS	STATUS
1	M/s. Muslim Trading Agencies	- All the items quoted by the said company are according to required specifications except item no. 142.	Evaluation details on page no. 1 and 7
2	M/s. Bio Trading	Samples Demanded and Evaluated Item No	Item No. 53-113, 117,133,135,141, were evaluated and found not satisfactory. Item No. 50,51,52,114,115,116,123,124,125,126,129,130,134,137,138,139,150,157,158,161,162,163 found satisfactory and are Acceptable
3	M/s. Moon Enterprises	Samples Demanded and Evaluated: Item No. 41,42,134,152,153,154,155,157,158,163,, 164,173,	Please refer to Page No. 3 and 4 for details of evaluation of item no. 152-155 and 41-42. Item No. 34, 157,158,163,173 are satisfactory. Item No 134, 164 Rejected.
4	M/s. Intexim Enterprises	- All the items quoted by the said company are according to required specifications	Please refer to Page No. 7 for details of evaluation of some items.
5	M/S.Three Star Medical	Samples Demanded and Evaluated: Items No. No.41,42,43,44,45,46,53,54,56,58,59,60,61-113 and 119,120,121,122,129,130,134,135,137,138, 141,142,144-149, 152,153,154,155.	Please refer to Page No1, 3, 4, and 7 for details of evaluation for some items. Item No. 32,33,34,47,48,119, 120,121,122,134,135,139,140,141,150, 151,156,157,158,159,160,163,170,173 are satisfactory. Item No,s 53-113, 129,130,137,138 are not satisfactory
6	M/s. Global Marketing	Samples Demanded and Evaluated: Item No. 44,45,46,53-106,	No Samples Provided. Hence all items rejected except item No.

Muhammad
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR

Muhammad

Fauzan
Dr. Abdul Waheed Khalid
DCP, M. PHD., FJPS (Pathology)
CONSULTANT PATHOLOGIST

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

			50,51,52 already evaluated and are satisfactory.
7	M/s. AS Enterprises	All the items are according to required specifications. <u>Samples Required</u> Item No .41,42,	Please refer to Page No.4, for details of evaluation for some items.
8	M/s. Alfa Scientific	Samples Demanded and Evaluated Item No 41, 42,142,152,154	Please refer to Page No.1,3,4 for details of evaluation for these items.
9	M/s. Musaji Adam & Sons	All the items are according to required specifications. Samples Demanded and Evaluated Item No.53,65-71,134,135,,137,138,139,140,141	Item No.53--134,137,138,139,140,157,161, 162 and 163 are satisfactory. Item No.135,141 are not satisfactory. Item No 150,151 Not demanded.
10	M/s. Roche Pakistan	All the quoted Items are according to required specifications. <u>Samples Required Item No 142</u>	Please refer to Page No. 1, for details of evaluation for some items. Item No 38,39,40 are already in use and satisfactory.
11	M/s. Sind Medical Store	<u>Samples Required ---</u> Items Nos. 41,42.	_ Please refer to Page No. 4 for details of evaluation for these items.
12	M/s. Mian Scientific	Not according to Required Specifications -Item No. 36-37 <u>43,44,45,46</u> <u>Samples Demanded</u> Item no 41,42,,134,152,153,154,155.	_ Please refer to Page No1,3,4,7 details of evaluation for some items. Item No. 36 and 37 not according to required specifications. 32,34,134,137,138, 156,157,158,163 are satisfactory.

[Signature]
 Additional Medical Superintendent
 Punjab Institute of Cardiology, LHR.

[Signature]
 Dr. Abdul Wadood Khalid
 DCP., M. Phil., FQPS (Pathology)
 CONSULTANT PATHOLOGIST
 Punjab Institute of Cardiology

DR. FAKHAR-UL-ISLAM
 Incharge Blood Bank
 Punjab Institute Of Cardiology
 Lahore.

	M/s. Popular International	- All the items quoted by the said company are according to required specifications	_ Please refer to Page No,7 for details of evaluation for some items. Items No 32, 47 are also satisfactory.
14	M/s. Hoorah Pharma	Not according to Required Specifications --- Item No. 38,39,40,44,45,46,142	_ Please refer to Page No 1,7 details of evaluation for some items. Item No 9,10,11,16,150,151 are according to required specifications. Item No 38,39,40 rejected as the brand is not demanded.
15	M/s. Zedco	Not according to Required Specifications Item No,144-149 <u>Samples Required</u> Item no 41,42,	_ Please refer to Page No,4, 7 details of evaluation for some items. Item no 144-149 Brand Not demanded. Hence Rejected
16	M/s. Biocept	Not according to Required Specifications. Item No,36,37,44,131,132.	_ Please refer to Page No7 details of evaluation for some items. item no 36,37 Rejected due to non-availability of distribution certificate from parent company. The offer for maintenance of instruments not as was required.
17	M/s. Rawal Bio Medical Service	No items quoted.	Refer to page no. 2 and 8
18	M/s. Basic Pharmaceuticals	The quoted item was not Demanded	
19	M/s. S. Ejazuddin & Co	- Both items quoted by the said company are according to required specifications.	No surety for maintenance of instruments attached as was required in tender. Hence rejected.

16
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.

Abdul Wadood Khan
Dr. Abdul Wadood Khan
DCP., M. Phil., FGPS (Pathology)
CONSULTANT PATHOLOGIST
Punjab Institute of Cardiology, Lahore.

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

20	M/s. Chemical House	Not according to Required Specifications 42, 166,167,168,169	Item No 30, 31 are acceptable. Items No. 42 rejected as this is a microwell assay. 166,167,168,169. Most of the EQA parameters of this program are with monthly samples. Our need is to have majority of samples biweekly in order to monitor the highly sensitive tests of cardiac patients. Hence this offer is rejected.
----	---------------------	--	--

Dated: 30-07-2016

Fakhar

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

Abdul Wadood Khalid
Dr. Abdul Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore

AS-11
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.

Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items
Final Evaluation Report of Samples of Anti HCV, HBs-Ag, HIV Devices And Other items
Blood Bank

Sr #	FIRMS NAME	Item No.	Make	Comments
1	M/s. Moon Enterprises	174Antiserum(Full Set)A,B,D Grouping 175-Bovine Albumin 22% 176-AntiHuman Globin 177 – Anti-HCV 178 – HBsAg 179-HIV devices 180-Plastic Test Tube With Green Cap 181-ICT Malaria Devices 182-ICT Syphilis Devices 183-Glass Slides	Imumed Imumed Imumed	Accepted Not Satisfactory Not-Satisfactory Accept Not Satisfactory Not Satisfactory Not Satisfactory
2	M/s. Three Star Medical Supplies	174Antiserum(Full Set)A,B,D Grouping 175-Bovine Albumin 22% 176-AntiHuman Globin 177 – Anti-HCV 178 – HBsAg 179-HIV devices	SBIO Fastep Fastep Fastep	Accepted No WHO Certificate – Rejected No WHO Certificate – Rejected No WHO Certificate – Rejected

[Signature]
 Addl. Medical Superintendent
 Punjab Institute of Cardiology, LHR.

[Signature]
Dr. Abdul Wadood Khalid
 DCP., M. Phil., FQPS (Pathology)
 CONSULTANT PATHOLOGIST
 Punjab Institute of Cardiology, Lahore

[Signature]
DR. FAKHAR-UL-ISLAM
 Incharge Blood Bank
 Punjab Institute Of Cardiology
 Lahore.

		180-Plastic Test Tube With Green Cap 181-ICT Malaria Devices 182-ICT Syphilis Devices 183-Glass Slides	China	Accepted No WHO Certificate – Rejected No WHO Certificate – Rejected Satisfactory
3	M/s.A.S.Enterprises	177 – Anti-HCV 178 – HBsAg 179-HIV devices 181-ICT Malaria Devices 182-ICT Syphilis Devices	SD Bioline Korea SD Bioline Korea SD Bioline Korea ABON ABON	Satisfactory Satisfactory Satisfactory Satisfactory Satisfactory
4	M/s. Alfa Scientific Store	177 – Anti-HCV 178 – HBsAg 179-HIV devices 181-ICT Malaria Devices 182-ICT Syphilis Devices	Bio Check Bio Check Bio Check Bio Check Bio Check	No WHO Certificate – Rejected No WHO Certificate – Rejected No WHO Certificate – Rejected No WHO Certificate – Rejected No WHO Certificate – Rejected
5	M/s.Sind Medical Store	174Antiserum(Full Set)A,B,D Grouping 175-Bovine Albumin 176-AntiHuman	DiaMed	Brand Not demanded

Muhammad
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.

Abdul Wadood Khalid
MCR, M. Phil, F&PS (Pathology)
CONSULTANT PATHOLOGIST
Punjab Institute of Cardiology, Lahore.

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

		Globin 177 – Anti-HCV 178 – HBsAg 179-HIV devices 181-JCT Malaria Devices 182-ICT Syphilis Devices	Intec PRC Intec PR	Not Satisfactory Rejected Not Satisfactory Rejected Not Satisfactory Rejected Not Satisfactory Rejected
6	M/s. Mian Scientific	179-HIV devices 180-Plastic Test Tube With Green Cap 181-ICT Malaria Devices 182-ICT Syphilis Devices 183-Glass Slides	CTK China CTK Boltan	No WHO Certificate – Rejected Ok No WHO Certificate – Rejected No WHO Certificate – Rejected No Sample
7	M/S Zedco	177 – Anti-HCV 178 – HBsAg 179-HIV devices 182-ICT Syphilis Devices	Mp Diagnostic Alere japan	No Sample No Sample No sample

* In lower titers of antibody, the risk of infection transmission exists if the device is non-responsive. As this is a screening method, it should include all borderline cases with no False Negatives.

Dated: 30-07-2016

Mi
 08.16
 Additional Medical Superintendent
 Punjab Institute of Cardiology, LHR.

Abdul
 Dr. Abdul Wadood Khalid
 DCP., M. Phil., FGPS (Pathology)
 CONSULTANT PATHOLOGIST
 Punjab Institute of Cardiology, Lahore

Fakhar
 Dr. FAKHAR-UL-ISLAM
 Incharge Blood Bank
 Punjab Institute Of Cardiology
 Lahore

Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items

Evaluation Report of Samples of Urine Strips (Item No. 142)

Sr. #	FIRM NAME	Make	Comments
1	M/s. Muslim Trading Agency	Meditest	Sample Tested. Performance Not Satisfactory.
2	M/s. Three Star	DFI Korea ComboStik	Brand Not demanded
3	M/S Alfa Scientific Store	Analyticon	Sample Tested. Performance Not Satisfactory.
4	M/s. Roche Pakistan	Combur 10	Sample Tested. Performance Satisfactory
5	M/S Mian Scientific	Uricon /Unimed	Brand Not Demanded
6	M/s. Hoorah Pharma	Siemens Health Care	Brand Not Demanded

* The details of evaluation report can be checked with the undersigned.

Dated: 30-07-2016

Fakhar

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

Abdul Wadood

Dr. Abdul Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore

Mi
09.08.16
Additional Medical Superintendent
Punjab Institute of Cardiology

Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items

Evaluation Report of Fully Automated Coagulation Analyzer to be placed on Reagent Rental Basis

Sr#	FIRMS NAME	Item No.	Make	Comments
1	M/s. Intexim Enterprises	PT& APTT	Precil Mindray C3100 China	Not According to Required Specifications.
2	M/s. Alfa Scientific Store	PT& APTT	Co R50	Not According to Required Specifications.
3	M/s. Zed Co	PT& APTT	Cp-3000 Abbott	Satisfactory and is according to Required Specifications
4	M/s. S.Fjazuddin & Co.	PT& APTT	SYSMEX	Satisfactory and is according to Required Specifications
5	Rawal Biomedical Services	PT& APTT	Thermo Scientific/ Technoclone	Not According to Required Specifications.

- The detail of evaluation report can be checked with the undersigned

Dated: 30-07-2016

Fakhar

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

Abdul Wadood

Dr. Abdul Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore

[Signature]
9-10-16

Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.


**TECHNICAL EVALUATION REPORT FOR PATHOLOGY LAB KITS, CHEMICALS
AND BLOOD BANK ITEMS.**

Evaluation Report of Provided Vacutainer Samples


Sr#	FIRM NAME	Item No.	Make	Comments
1	M/s. Moon Enterprises	152- EDTA	ImuMed	Satisfactory
		153 – Sodium Citrate	ImuMed	Satisfactory
		154 – Lithium Heparin	ImuMed	Satisfactory
		155 – Glass Slides	China	Not Satisfactory
2	M/s. Three Star	152- EDTA	Bio Vac	Satisfactory
		153 – Sodium Citrate	Bio Vac	Satisfactory
		154 – Lithium Heparin	Bio Vac	Satisfactory
		155 – Glass Slides	China	Satisfactory
3	M/s. Alfa Scientific	152- EDTA	Xinle	No Samples
		154 – Lithium Heparin	Xinle	No Samples
4	M/s. Mian Scientific	152- EDTA	Improve	Not Satisfactory
		153 – Sodium Citrate	Improve	Not Satisfactory
		154 – Lithium Heparin	Improve	Not Satisfactory
		155 – Glass Slide	Bolton	No Samples

- The details of evaluation report can be checked with the undersigned.


Dated: 30-07-2016



DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.



Dr. Abdul Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore.



Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.

Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items**Final Evaluation Report of Samples of Anti HCV and HBs-Ag Rapid Test Cassetts**

Sr #	FIRMS NAME	Item No.	Make	Comments
1	M/s. Moon Enterprises	41 – Anti-HCV 42 – HBsAg	Imumed Imumed	Performance Not Satisfactory -- Rejected Performance Not-Satisfactory - Rejected
2	M/s. Three Star Medical Supplies	41 – Anti-HCV 42 – HBsAg	Fastep Fastep	No WHO Certificate – Rejected No WHO Certificate – Rejected
3	M/s.A.S.Enterprises	41 – Anti-HCV 42 – HBsAg	SD Bioline Korea SD Bioline Korea	Performance Satisfactory --- Accepted Performance Satisfactory --- Accepted
4	M/s. Alfa Scientific Store	41 – Anti-HCV 42 – HBsAg	Bio Check Bio Check	No WHO Certificate – Rejected No WHO Certificate – Rejected
5	M/s.Sind Medical Store	41 – Anti-HCV 42 – HBsAg	Intec PRC Intec PR	Performance Not Satisfactory – Rejected Performance Not-Satisfactory - Rejected
6	M/s. Mian Scientific	41 – Anti-HCV 42 – HBsAg	CTK CTK	No WHO Certificate – Rejected No WHO Certificate – Rejected
7	M/S Zedco	41 – Anti-HCV 42 – HBsAg	MP Diagnostic Alec Japan	No Samples Provided No Sample Provided

Dated: 30-07-2016

Fakhar

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

Handwritten signature and date: 30-08-16
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.

Handwritten signature
Dr. Abdul Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore

Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items

Final Evaluation Report of 3-Part Differential Haematology Analyzer to be Placed on Reagent Rental Basis

Sr.#	FIRMS NAME	Item	Make	Comments
4	M/s. Seico Scientific Traders	3-Part Differential Hematology Analyzer	Diatron Hungary Abacus-300	Instrument Placed for Demonstration. Performance Satisfactory.
6	M/S. Sindh Medical Stores	3-Part Differential Hematology Analyzer	MS 4sc	Instrument Placed for Demonstration. Performance Satisfactory.
7	M/s. S. Ejazuddin & Co.	3-Part Differential Hematology Analyzer	Sysmex Kx-2I	Instrument Already Running. Performance Satisfactory

Dated: 30-07-2016

Faizan

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.

Abdu Wadood Khalid
Dr. Abdu. Wadood Khalid
Sr. Consultant Pathologist
PIC, Lahore

[Signature]
31.08.16

Additional Medical Superintendent
Punjab Institute of Cardiology, L.I.R.


Technical Evaluation Report of Lab Kits, Chemicals and Blood Bank Items
Final Evaluation Report of Samples of Serology Kits

Sr#	FIRMS NAME	Item No.	Make	Comments
1	M/s. Muslim Trading Agency	43 - ANA 44 - ASOT 45 - CRP 46 - RAF 144-149 Widal	Randox Immunostics	Satisfactory Satisfactory Satisfactory Satisfactory Satisfactory
2	M/s. Intexim Enterprises	43 - ANA 44 - ASOT 45 - CRP 46 - RAF 144-149 Widal	Spinreact Spinreact	Satisfactory Satisfactory Satisfactory Satisfactory Satisfactory
3	M/s. Three Star	43 - ANA 44 - ASOT 45 - CRP 46 - RAF 144-149 Widal	SBIO SBIO	Not Demanded Brand Hence Rejected
4	M/s. Global Marketing	44 - ASOT 45 - CRP 46 - RAF	Omega,UK	Not Demanded Brand Hence Rejected
5	M/s. Mian Scientific	43 - ANA 44 - ASOT 45 - CRP 46 - RAF	BioMed	Not Demanded Brand Hence Rejected
6	M/s. Popular International	43 - ANA 44 - ASOT 45 - CRP 46 - RAF	Human	Satisfactory Satisfactory Satisfactory Satisfactory
7	M/s. Hoorah Pharma	44 - ASOT 45 - CRP 46 - RAF	Siemens Health Care Diagnostic	Not Demanded Brand Hence Rejected
8	M/s. Zedco	144-149 Widal	Lorne UK	Not Demanded Brand Hence Rejected
9	M/s. Biocept	44 - ASOT 45 - CRP 46 - RAF	Randox	Letter of sole distribution for serology not there. Hence Rejected


Dated: 30-07-2016



DR. FAKHAR-UL-ISLAM
 Incharge Blood Bank
 Punjab Institute Of Cardiology
 Lahore.



Dr. Abdul Wadood Khalid
 Sr. Consultant Pathologist
 PIC, Lahore



Additional Medical Superintendent
 Punjab Institute of Cardiology, LHR.


TECHNICAL EVALUATION REPORT FOR PATHOLOGY LAB KITS, CHEMICALS AND BLOOD BANK ITEMS
EVALUATION REPORT – Troponin T, HBsAg, Anti HCV and TFTs by Chemilumiscence

Sr. #	FIRMS NAME	Item No.	Make	Comments
1	M/S. Roche Pakistan	Troponin-T T3 Free T4 Free TSH(HS) Anti HCV HbsAg	ELECSYS	- Offer is according to required specifications. Hence accepted.
2	M/s. Hooraa Pharma	T3 Free T4 Free TSH(HS) Anti Hcv HbsAg	Siemen /Advia Centure	Past experience of thyroid profile of same make in PIC was not satisfactory. Please get the kits of thyroid and B + C Elisa evaluated for satisfactory performance within one week.
3	M/S. Zedco	T3 Free T4 Free TSH(HS) Anti HCV HbsAg	Architect 1000 SR	Offer is according to required specifications. Hence accepted.
4	Rawal Biomedical Services	T3 Free T4 Free TSH(HS) Hs Troponin-I		Offer is only for thyroid profile. HbsAg and Anti HCV are not offered. Troponin – I was not demanded. Subdivision of samples on multiple instruments not feasible. Hence the bid is not acceptable.

Dated = 30-07-2016

Fakhar

DR. FAKHAR-UL-ISLAM
Incharge Blood Bank
Punjab Institute Of Cardiology
Lahore.



Dr. Abdul Waheed Khalid
Senior Consultant Pathologist,
PIC, Lahore


4-98-16
Additional Medical Superintendent
Punjab Institute of Cardiology, LHR.