

CSSD

PUNJAB INSTITUTE OF CARDIOLOGY, LAHORE.
 MEDICAL ACCESSORIES & SPARE PARTS, TENDER 2016 - 2017
 TECHNICAL EVALUATION REPORT


S.#.	Item Name	REQUIRED SIZE	SPECIFICATION	QTY.	M/S quintex	M/S Medijorge	M/S medisource	M/S TOTAL TECHNOLOGY	M/S CARDIAC CARE
					Quoted Brand, Origin	Quoted Brand, Origin	Quoted Brand, Origin	Quoted Brand, Origin	Quoted Brand, Origin
1	Steam Indicator Strips	Standard		20	cat ref/ CDD20 (class 4 EN ISO 11140-1) terragene argentina	PART CAT 2311AB albert browne UK Sctris corp usa	STERIS CROP USA	C-S-P-4SV1 gbe Germany	N/Q
Remarks					SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	SUBMITTED STRIP QUALITY IS POOR NOT ACCEPTABLE	
Technical Eligibility of product					NOT ELIGIBLE	Eligible	NOT ELIGIBLE	Not Eligible	
2	EO Indicator Strips	Standard		8	cat ref/ CD10 (class 4 EN ISO 11140-1) terragene Argentina	PART CAT 2420AB Albert Browne UK Sctris corp. usa	STERIS CROP USA	C-E-P-4-SV3 gbe Germany	N/Q
Remarks					SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	SUBMITTED STRIP QUALITY IS POOR NOT ACCEPTABLE	
Technical Eligibility of product					NOT ELIGIBLE	Eligible	NOT ELIGIBLE	Not Eligible	
3	Flat Paper Roll	3"		15	UNI EN ISO 9001-UNI EN ISO 13485 STANDERD E-line Srl Italy	NWFL 75200 Sctris corp usa EEC/ TURKEY	E-LINE ITALY	PMS TURKEY	E-LINE ITALY
Remarks					SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	Sample submitted Acceptable
Technical Eligibility of product					NOT ELIGIBLE	Eligible	NOT ELIGIBLE	Eligible	Eligible

64-08-16

ESSD

4	Flat Paper Roll	4"	Safe sterilizing packing roll. Suitable for steam sterilization and Ethylen Oxide sterilization. Free of toxic material and unpleasant smell. Weight 70gm. Transparent. Pinhole free. Heat sealable. Chemicals indicator printed in the Sealed Layer.	15	UNI EN ISO 9001-UNI EN ISO 13485 STANDERD E-line Srl Italy	NWFL 10200 Sertis corp usa EEC/ TURKEY	PMS TURKEY	E-LINE ITALY
		Remarks		SAMPLE NOT SUBMITTED NOT ACCEPTABLE		Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable
Technical Eligibility of product								
5	Flat Paper Roll	6"	Same as items No. 04	20	UNI EN ISO 9001-UNI EN ISO 13485 STANDERD E-line Srl Italy	Eligible NWFL 15200 Sertis corp usa EEC/ TURKEY	PMS TURKEY	Eligible E-LINE ITALY
		Remarks		SAMPLE NOT SUBMITTED NOT ACCEPTABLE		Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable
Technical Eligibility of product								
6	Flat Paper Roll	8"	Same as items No. 04	20	UNI EN ISO 9001-UNI EN ISO 13485 STANDERD E-line Srl Italy	Eligible NWFL 20200 Sertis corp usa EEC/ TURKEY	PMS TURKEY	Eligible E-LINE ITALY
		Remarks		SAMPLE NOT SUBMITTED NOT ACCEPTABLE		Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable
Technical Eligibility of product								
7	Flat Paper Roll	10"	Same as items No. 04	15	N.Q	Eligible NWFL 25200 Sertis corp usa EEC/ TURKEY	PMS TURKEY	Eligible E-LINE ITALY
Technical Eligibility of product								

ESSD

Technical Eligibility of product			Remarks	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	Sample submitted Acceptable	
8	Flat Paper Roll	12"	Same as items No. 04	15	UNI EN ISO 9001-UNI EN ISO 13485 STANDERD E-line Srl Italy	Eligible NWFL 30200 Sectris corp usa EEC/ TURKEY	NOT ELIGIBLE E-LINE ITALY PMS TURKEY	Eligible E-LINE ITALY
Technical Eligibility of product			Remarks	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	Sample submitted Acceptable
9	Flat Paper Roll	16"	Same as items No. 04	6	N.Q.	Eligible NWFL 40200 Sectris corp usa EEC/ TURKEY	NOT ELIGIBLE E-LINE ITALY	Eligible E-LINE ITALY
Technical Eligibility of product			Remarks	NOT ELIGIBLE	Sample submitted Acceptable	SAMPLE NOT SUBMITTED NOT ACCEPTABLE	Sample submitted Acceptable	Sample submitted Acceptable
10	Flat Paper Pouch	6"		40	N.Q.	Eligible NWFP 1520 Sectris corp usa EEC/ TURKEY OPTON A NWFP 1525 Sectris corp usa EEC/ TURKEY OPTON B NWFP 1530 Sectris corp usa EEC/ TURKEY	NOT ELIGIBLE E-LINE ITALY STERIS CORP USA	Eligible E-LINE ITALY N.Q.
Technical Eligibility of product			Remarks			MULTIPLE OPTON NOT ACCEPTABLE NOT ELIGIBLE	SAMPLE NOT SUBMITTED NOT ACCEPTABLE NOT ELIGIBLE	Sample not submitted Not Acceptable Not Eligible

ESSD

11	Flat Paper Pouch	8"	30	N/Q	NWFP 2040 Steris corp usa EEC/TURKEY OPTION A NWFP 2060 Steris corp usa EEC/TURKEY	STERIS CORP USA	N/Q	E-LINE ITALY
Remarks								
Technical Eligibility of product								
12	Flat Paper Pouch	10"	15	N/Q	NOT ELIGIBLE NWFP 1235 Steris corp usa EEC/TURKEY OPTION A NWFP 2550 Steris corp usa EEC/TURKEY	SIGMA TAIWAN	N/Q	E-LINE ITALY
Remarks								
Technical Eligibility of product								
13	EO Gas Cartridge with Accessories	EO Sterilizer	500	N/Q	NOT ELIGIBLE	N/Q	ANDERSON UK	N/Q
Remarks								
Technical Eligibility of product								

NOTE: The firms offer will be technically graded only after approval of samples. The offers without samples will not be entertained.

Sample not submitted
Not Acceptable

Sample not submitted
Not Acceptable

Not Eligible

Acceptable

Eligible

CSSD CONSUMABLE REVIEW COMMENTS


M/s Quintex Medical

M/s Quintex Medical requested that the firm has submitted samples within the given time frame but it was not considered by technical committee. The samples submitted by M/s Quintex Medical and found it as per specification and hence technically qualified.

M/s Total Technologies

M/s Total Technologies submitted that their quoted steam indicator strip and EO Indicator strip have been technically rejected. However our quoted strips are of GKE, Germany

The End user re-evaluated the firm samples and the quality certificate documentation, provided by the firm was as per requirement. Hence firm offer for steam Indicator strip and EO Indicator Strip stands technically qualified.



End User/ Senior Technician of Respective Depart
PIC, Lahore